

# REFEX pH electrode hard wiring to ENDRESS + HAUSER LIQUILINE CM42 pH Meter

Wiring Connections WITHOUT Liquid Earth/Solution Ground:

**Clear wire** (pH glass) term 317

**Braid** (screen) and **Green wire** (Ref) term 320

**White wire** Pt1000/Pt100 term 113

**Red wire** Pt1000/Pt100 term 112

**Install jumper between terminals 111 and 113**



## Important:

To Change Configuration for Liquid Earth/Solution Ground

SELECT - SETUP

SELECT - SENSOR PH/ORP

SCROLL DOWN TO - POTENTIAL MATCHING

SCROLL DOWN TO - DIAGNOSTICS LIMITS

SELECT EITHER - WITH POTENTIAL MATCHING

or

WITHOUT POTENTIAL MATCHING

